

## VERIFICATION OF COMPLIANCE

**Issue Date:** May 6, 2024  
**Applicant:** FSP Group Inc.  
**Address:** No 22, Jianguo East Road, Taoyuan City, Taiwan 330, R.O.C.  
**Manufacturer:** FSP Group Inc.  
**Address:** No 22, Jianguo East Road, Taoyuan City, Taiwan 330, R.O.C.  
**Product:** Switching Power Adapter  
**Brand Name/Trade Mark:** FSP  
**Model/Type:** FSP065M-DUA; FSP045M-DUA; FSP065M-DUB; FSP045M-DUB; FSP045M-DUAXXX ("X"=0-9, A-Z, hyphen or blank.); FSP045M-DUBXXX ("X"=0-9, A-Z, hyphen or blank.); FSP065M-DUAXXX ("X"=0-9, A-Z, hyphen or blank.); FSP065M-DUBXXX ("X"=0-9, A-Z, hyphen or blank.)  
**Applicable Standards:** EN 60601-1-2: 2015 + A1: 2021  
IEC 60601-1-2: 2014 + A1: 2020  
BS EN 60601-1-2: 2015 + A1: 2021  
CISPR 11: 2015 + A1: 2016 + A2: 2019 (Ed 6.2)  
IEC 61000-3-2: 2018 + A1: 2020 (Ed. 5.1)  
IEC 61000-3-3: 2013 + AMD1: 2017 + AMD2: 2021 CSV (Ed. 3.2)  
IEC 61000-4-2: 2008  
IEC 61000-4-3: 2020 (Ed. 4.0)  
IEC 61000-4-4: 2012  
IEC 61000-4-5: 2014 + A1: 2017  
IEC 61000-4-6: 2013 + COR1: 2015  
IEC 61000-4-8: 2009  
IEC 61000-4-11: 2020 + COR1: 2020 + COR2: 2022 (Ed. 3.0)  
IEC 60601-1-2: 2014 + A1: 2020 Subclause 8.11 (IEC 61000-4-39 for §8.10)  
**Test Laboratory:** Compliance Certification Services Inc.  
Xindian Lab  
No. 163-1, Jhongsheng Rd., Xindian Dist., New Taipei City, Taiwan  
**Test Report No.:** TMXD2404001290DE, dated on May 6, 2024  
**Conclusion:** Based upon a review of the Test Report(s), the tested sample of the product mentioned above is deemed to comply with the requirements of the above standards.  
**Note:** This verification is only valid for the product and configuration described and in conjunction with the test report as detailed above.

**Authorised Signatory:**



**Compliance Certification Services Inc.**  
**Jason Lee**  
**Section Manager**